



HERMON LABORATORIES

CERTIFICATE

CERTIFICATE OF CONFORMITY

With EN 60601-1: 2006 + A11:2011 + A1:2013; IEC 60601 1:2005 (Third Edition) + CORR. 1:2006 + CORR. 2:2007 + A1:2012 (or IEC 60601-1: 2012 reprint); ANSI/AAMI ES60601-1:2005/(R):2012 +A1:2012+ C1:2009+ A2:2010; CAN/CSA-C22.2 No. 60601-1:14 standards

Certificate Number NOVSAF_EN.41289C

This certificate of conformity has been granted to the applicant based on the results of tests and evaluations, performed by Hermon Laboratories from 11 May to 11 August, 2015, 26 July, 2020 and 23 December, 2020 on representative sample of the specified product.

Product description

Tested item:	Cancer Treatment Device in the brain
Model:	Optune TFH9100 (NovoTTF-200A)
Serial number:	TFH207885
Hardware version:	For EUT: A2, For RDL communication cable: A
Software release:	For EUT: 40.1.10.1, For NovoTerminal: 1.0.4

Applicant details

Name:	NovoCure Ltd.
Address:	Topaz Building, MATAM Center, Haifa 3190500, Israel

Manufacturer details

Name:	NovoCure GmbH
Address:	Park 6 / Platz 10, 6039 Root D4, Switzerland

This is to certify that the tested product sample satisfies the requirements of the above listed standard/s.

Measurement/test results are contained in the test report: NOVSAF_EN.41289.

The comments in the associated test report shall be taken into account and used in conjunction with this certificate

Michael Brun,
Product Safety Group Manager
Hermon Laboratories Ltd.

December 31, 2020

Page 1 of 1

EXPERTS IN GLOBAL COMPLIANCE SOLUTIONS



EMC



Radio



Telecom



Environmental



Product
Safety



International
Approvals

Hermon Laboratories Ltd.

Hatachana St., POB 23, Binyamina 30500 Israel

Phone: +972 4 628 8001, Fax: +972 4 628 8277

Email: mail@hermonlabs.com, www.hermonlabs.com