



HERMON LABORATORIES

CERTIFICATE OF CONFORMITY

With EN 60601-1:2006 + A11:2011 + A1:2013 + A2:2021; IEC 60601-1:2005 + A1:2012 + A2:2020; ANSI/AAMI ES60601-1:2005/(R)2012 + A1:2012+ C1:2009/(R)2012+ A2:2010/(R)2012; CAN/CSA-C22.2 No. 60601-1:14 standards

Certificate Number NOVSAF_EN.144249C

This certificate of conformity has been granted to the applicant based on the results of tests and evaluations, performed by Hermon Laboratories 25 December, 2018 - 6 October, 2019; 23 July, 2020; 4 January – 26 June, 2023; 20 November – 28 November, 2023; 31 March – 01 April, 2025, 5 – 9 June, 2025 on representative sample of the specified product.

Product description

Tested item: Cancer Treatment Device in Torso area
Models: Optune Lua (NovoTTF-200T), model: TFT9200
Optune Pax (NovoTTF-200P), model: TFP9200
Serial number: 001
Hardware version: TFT1200_BOM_NVC A01
Software release: 40.1.10.1

Applicant details

Name: Novocure Ltd.
Address: Topaz Building, MATAM Center, Haifa 3190500, Israel

Manufacturer details

Name: Novocure GmbH
Address: Park 6 / Platz 10, 6039 Root, Business Village D4, Switzerland

This is to certify that the tested product sample satisfies the requirements of the above listed standard/s.

Measurement/test results are contained in the test report: NOVSAF_EN.144249.
The comments in the associated test report shall be taken into account and used in conjunction with this certificate

Ilan Benihass,
Product Safety Head of Department
Hermon Laboratories Ltd.

26 June, 2025

Page 1 of 1

EXPERTS IN GLOBAL COMPLIANCE SOLUTIONS



EMC



Radio



Telecom



Environmental



Product
Safety



International
Approvals

Hermon Laboratories Ltd.

Hatachana St., POB 23, Binyamina 30500 Israel

Phone: +972 4 628 8001, Fax: +972 4 628 8277

Email: mail@hermonlabs.com, www.hermonlabs.com